

**NORTH AMERICAN
POWER CORDS**



Hospital Grade Power Cords

QUALTEK ELECTRONICS


Hospital Grade - Unshielded with North American Color Code



I



J



K



L





Part Number	Figure	Length	Cordage	# Cond. AWG	Amp Rating	Plug No.	Conn No.
232002-06	I	8' 0"	SJT+	3 x 18	10	019*	--
232003-06	I	10' 0"	SJT+	3 x 18	10	019*	--
232010-06	I	8' 0"	SJT+	3 x 18	10	019**	--
232011-06	I	10' 0"	SJT+	3 x 18	10	019**	--
233002-06	J	8' 0"	SJT+	3 x 18	10	019*	005N
233003-06	J	10' 0"	SJT+	3 x 18	10	019*	005N
233008-06	J	8' 0"	SJT+	3 x 18	10	019**	005N
233009-06	J	10' 0"	SJT+	3 x 18	10	019**	005N
242003-06	K	10' 0"	SJT+	3 x 18	10	019L*	--
243003-06	L	10' 0"	SJT+	3 x 18	10	019L*	005N

- * Hospital Grade Clear Plug
- ** Hospital Grade Gray Plug
- + North American Color Code (Black/White/Green)

All Cords have 60°C and 125VAC unless marked as above

 [Home](#) /  [Products](#) /  [Literature Request](#) /  [What's NEW](#)

Qualtek Electronics *Meeting Global Demands with Design Solutions*

7675 Jenther Drive  Mentor, Ohio 44060  (888) 258-3468  (440) 951-3300  FAX: (440) 951-7252

E-Mail mailbox@qualtekusa.com  Web: www.qualtekusa.com

Copyright © 2004 Qualtek Electronics Corporation. All rights reserved.

Updated: March, 2004