

	TENMA	DATA SHEET (PAGE 1 OF 2).	GLOSSARY :
Black : 76-1652 Red : 76-1554			ACCESSIBLE. Able to be touched with a standard test finger or test pin BASIC INSULATION. Insulation of HAZARDOUS LIVE parts which provides basic protection.
			CAT II. Measurement or overvoltage category II. For measurement per- formed on / equipment connected to the building wiring.
			CAT III. Measurement or overvoltage category III. For measurement performed on / equipment connected to part of a building wiring installa
			CAT IV. Measurement or overvoltage category IV. For measurement performed on / equipment connected to the origin of the electrical supp building.
			CLEARANCE. Shortest distance in air between two conductive parts.
			CREEPAGE DISTANCE. Shortest distance along the surface of a soli insulating material between two conductive parts.
			CTI. Comparative Tracking Index of the insulating material in accorda with IEC 60112.
			DOUBLE INSULATION. Insulation comprising both BASIC INSUL/ and SUPPLEMENTARY INSULATION.
	Electrical safety	Very low voltages only :	EN / IEC 60529:2001. The 2001 version of the European / international standard regarding the degrees of protection provided by enclosures.
	33 V AC / 70 V DC	33 Ý AC / 70 ϔ DC, 36 A (at +40 °C).	EN / IEC 61010-1:2010. The latest version (in February 2012) of the European / international standard regarding the safety requirements for electrical equipment for measurement, control, and laboratory use – Pr General requirements. Version year 2010.
	Operating temperature range	e -20 °C mini., +80 °C maxi. (please see above too).	EN / IEC 61010-031:2008. The latest version (in February 2012) of th European / international standard regarding the safety requirements for electrical equipment for measurement, control and laboratory use – Pa Safety requirements for hand-held probe assemblies for electrical mea
	Conformity	 European Directive "RoHS" 2011/65/EU. European REACH regulation n°1907 / 2006. 	ment and test. Version year 2008. "LVD". European Directive 2006/95/EC on the harmonisation of the law Member States relating to electrical equipment designed for use within certain voltage limits. (Usually called the Low Voltage Directive.)
	Environment	 "RoHS" compliant, Pb ≤ 4 % in conductor, Pb ≤ 0.1 % in insulator, Hg ≤ 0.1 %, Cr VI ≤ 0.1 %, Cd ≤ 0.01 %, PBB ≤ 0.1 %, and PBDE ≤ 0.1 %. REACH compliant, no substances from the candidate list of SVHC for authorisation at mass 	MAINS. Low-voltage electricity supply system to which the equipmen concerned is designed to be connected for the purpose of powering the equipment.
		concentrations greater than 0.1 %.	MAINS CIRCUIT. Circuit which is intended to be directly connected MAINS for the purpose of powering the equipment.
	Materials	Conductor : nickel-coated brass, brass, steel. Insulators and lantern contact spring, please contact us	OVERVOLTAGE CATEGORY. Numeral defining a TRANSIENT O VOLTAGE condition.
			POLLUTION. Addition of foreign matter, solid, liquid or gaseous (ion gases), that may produce a reduction of dielectric strength or surface resistivity.
	Colors	Black Red	POLLUTION DEGREE. Numeral indicating the level of POLLUTION th may be present in the environment. POLLUTION DEGREE 1. No POLLUTION or only dry, non-conductive POLLUTION DEGREE 2. Only non-conductive POLLUTION occurs exe that occasionally a temporary conductivity caused by condensation is expected. REINFORCED INSULATION. Insulation which provides protection again
	Weight	0.015 kg.	
	Origin	Designed and manufactured in France.	
	Reliability benchmark	Year of 1st placing on the market 1993.	
	Packaging	One piece per bag.	REINFORCED INSULATION. Insulation which provides protection a electric shock not less than that provided by DOUBLE INSULATION "RoHS". European Directive 2011/65/EU on the restriction of the use ' certain hazardous substances in electrical and electronic equipment.
			SOLID INSULATION. Insulating materials.
			SUPPLEMENTARY INSULATION. Independent insulation applied addition to BASIC INSULATION in order to provide protection against electric shock in the event of a failure of BASIC INSULATIC
			TRANSIENT OVERVOLTAGE. Short duration overvoltage of a few milliseconds or less, oscillatory or non-oscillatory, usually highly dar

Fichier « DataSheet_TENMA_1089 ». 21/10/2016. Indice A. Information in this publication supersedes all earlier versions. Specifications subject to change without notice. © 2016.

WORKING VOLTAGE. Highest r.m.s. value of the a.c. or d.c. voltage across any particular insulation which can occur when the equipment is supplied at rated voltage.